



Apple App Store



Motion Intelligence 17+
DJO, LLC
Designed for iPad
★★★★★ 5.0 • 3 Ratings
Free


Android Google Play Store




Motion Intelligence
DJO Global Medical
Everyone
Add to Wishlist

Mobile App Walk-Through


1 Enter your pin to login.




2 Follow the prompts to connect your X4 brace.



3




4




5

Once complete, tap "Next" to go to main dashboard.




5

Tap "Continue" to start your exercises.




6

Tap the black bar at the bottom to start.




7

Follow the prompts to complete your exercises.




8

Tap "Next" to return to main dashboard.



Brace Failed to Connect Automatically

If you receive the error message to the right, tap the “OK” button, then grab your X4 brace and press the clear button. It will turn blue to reconnect.



Follow these steps if you continue to receive the error to unpair your X4 brace and reconnect:

- 1 Tap “OK” to clear the message
- 2 Tap “Settings” (located in the bottom navigation menu) then, “Forget Brace”
- 3 Navigate to your phone’s Bluetooth settings, select your X4 brace, and tap “Forget this device”
- 4 Uninstall the MI app
- 5 Navigate to your app store and reinstall the MI app
- 6 After download is complete, open the MI app
- 7 Open your original text message received the day you received your X4 brace. Tap on the link, which will take you back to the MI app
- 8 Follow the on-screen prompts and press the clear button on your X4 brace when the app states it is trying to find/pair your X4 brace